



Audix Technology (Shanghai) Co., Ltd.

3F 34Bldg 680 Guiping Rd.,
Caohejing Hi-Tech Park, Shanghai, China 200233

Tel: +86-21-64955500
Fax: +86-21-64955491
E-Mail: audixaci@audix.com

No. ACI-E000133

Statement of Compliance

Regarding the statement of products which are in the scope of
the Council Directive 89/336/EEC, 92/31/EEC,

Shanghai Jenco Electronics, Ltd.

18 Wang Dong Zhong Road, Sijing Town, Songjiang Shanghai China
has successfully demonstrated that its product

Conductivity Meter, pH Meter, Dissolved Oxygen Meter
(M/N: 3173R, 6173R, 9173R)
(S/N: E04021301, E03081402, E03112302)


is in compliance with

EN61326:2002 (CISPR 11:1999
IEC 61000-3-2:2001, IEC 61000-3-3:1995,
IEC 61000-4-2:2001, IEC 61000-4-3:2002,
IEC 61000-4-4:1995/A1:2000/A2:2001,
IEC 61000-4-5:2001, IEC 61000-4-6:2001,
IEC 61000-4-11:2001)

as described in the Test Report ACI-E03052

(Date: Feb 2002)  For and on behalf of
Audix Technology (Shanghai) Co., Ltd.




Authorized Signatory **JASON CHEN**
Deputy Manager

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.